

.

From: Piro, Peter (DPH)
Sent: Friday, March 09, 2012 8:50 AM
To: Saunders, Della (DPH); Corbett, Kate (DPH); Glazer, Lisa (DPH); Frasca, Daniela (DPH); Renczkowski, Daniel (DPH); Medina, Nicole (DPH); Lawler, Michael (DPH); Lleshi, Hevis (DPH); Piro, Peter (DPH)
Cc: Nassif, Julianne (DPH); Salemi, Charles (DPH)
Subject: GC/MS QC Procedure Update

If a GC/MS run is discarded, continue the usual documentation on the GC/MS Daily QC Check sheet and the GC/MS Daily Injector/Column Check sheet. In addition, place an * next to Run Discarded and in the comments section record the sequence with a very brief explanation of why the run was discarded. Thank you in advance.

Peter Piro
GC/MS Laboratory Supervisor I
Department of Public Health
Hinton State Laboratory Institute
Forensic Drug Laboratory, Rm 361
305 South Street
Jamaica Plain, Ma 02130
617-983-6634
Peter.Piro@state.ma.us